

## Review Article

# Forty Years of Laser-Induced Breakdown Spectroscopy and Laser and Particle Beams

Vincenzo Palleschi 

*Applied and Laser Spectroscopy Laboratory, Institute of Chemistry of Organometallic Compounds,  
National Research Council Research Area, Via G. Moruzzi 1-56124, Pisa, Italy*

Correspondence should be addressed to Vincenzo Palleschi; [vincenzo.palleschi@cnr.it](mailto:vincenzo.palleschi@cnr.it)

Received 15 September 2022; Revised 24 May 2023; Accepted 7 June 2023; Published 19 June 2023

Academic Editor: Bhuvanesh Ramakrishna

Copyright © 2023 Vincenzo Palleschi. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

The laser-induced breakdown spectroscopy (LIBS) technique is one of the most promising laser-based analytical techniques. Coincidentally, the LIBS acronym was proposed by Radziemski and Loree in two seminal papers published in 1981, almost at the same time in which the Laser and Particle Beams journal started its publication. In this contribution, the evolution of the LIBS technique is discussed following a chronological collection of key papers in LIBS, some of which were in fact published on LPB.

## 1. Introduction

Laser-induced breakdown spectroscopy (LIBS) is an atomic emission spectroscopic technique based on the spectral analysis of the plasma induced by a pulsed laser beam in gas, liquid, or solid targets to obtain information about the materials under study.

The principles of the LIBS technique are deeply rooted on the preexisting knowledge on flame and plasma spectroscopy which largely precedes the discovery of the laser. After the introduction of the laser, techniques similar to modern LIBS were proposed [1]; however, as in the 1962 Brech and Cross paper, at those times, the laser was used essentially for ablating a solid sample, while the excitation of the material was obtained through an electrical spark. The main characteristics of LIBS as it is now practiced are, on the other hand, to use the laser for obtaining at the same time the sampling of the material and its heating for producing the atomic optical emission [2].

This characteristic is peculiar of the LIBS technique and brings the exceptional advantages of operating on not treated materials in very short time, which in turn allow the use of the technique for remote in situ analysis in hostile environments. On the other hand, the use of a single tool for sampling and excitation prevents the possibility of

independent optimization of the two processes, leading to analytical performances that are usually considered modest with respect to other conventional laboratory spectrochemical techniques.

A typical LIBS experiment involves the use of a pulsed laser, typically Nd:YAG, at the fundamental wavelength of 1064 nm, emitting pulses of a few nanoseconds with energy of several tens of milli-Joules and maximum repetition rates of a few Hertz. The light emitted by the laser-induced plasma is collected and sent to a spectrometer (gated or ungated, narrow- or wideband) where the signal is analysed using a suitable delay after the laser pulse, to reduce the continuum bremsstrahlung emission (see Figure 1).

Many alternative experimental configurations have been realized, though. A description of some of them is given in [3].

A typical LIBS spectrum, acquired with a broadband spectrometer, is shown in Figure 2.

The attribution of the emission lines to the corresponding atomic species is usually done manually, based on the information contained in the NIST database of atomic lines [4], although methods for automatic identification of the lines were also proposed [5].

In the following, we will discuss the exceptional evolution of the LIBS technique in the last 40 years, also




FIGURE 1: A typical LIBS setup.




FIGURE 2: A typical broadband LIBS spectrum (steel).

highlighting the important role that the *Laser and Particle Beams* journal has had in this evolution.

## 2. 1981–1990: The Early Years

The first papers where the acronym LIBS was originally proposed were published by Radziemski and Loree [6, 7] in 1981. The two authors, researchers of the Los Alamos National Laboratory (Los Alamos, New Mexico, USA), outlined the principle of the LIBS technique in two companion papers, the first dealing with time-integrated detection of the plasma emission, the second discussing the analytical advantages of a time-resolved detection. The authors analysed by LIBS sodium and potassium in a coal gasifier product, airborne beryllium and phosphorous, sulphur, fluorine, and chlorine (the latter three elements particularly complex to detect by LIBS) in atmosphere. For the next 10 years, the research on LIBS remained essentially confined in North America; with the arrival of the Radziemski group of David

Cremers, the application of the LIBS technique was extended to many other interesting fields, such as the study of aerosols [8], the analysis of liquids [9], detection of beryllium in air [10] and in beryllium-copper alloys [11], and detection of uranium in solution [12] and cadmium, lead, and zinc in aerosol [13].

## 3. 1991–2000: Evolution of LIBS

In 1991, the Pisa group published a paper dealing with the quantitative determination of pollutants in air by LIBS [14]. The paper was published on *Laser and Particle Beams*, and it represented the first work on LIBS published by a group outside the USA.

In the following years, other works on LIBS were published in Europe (determination of carbon in steel by the Spanish group of Aragón et al. [15, 16]) and in Canada (quantitative analysis of aluminium alloys [17]).

In the 1990–2000 decade, several papers on LIBS were published using the “LIPS” (laser-induced plasma spectroscopy) acronym. This occurred mostly in Europe (see, for example [18]), but some groups in Canada [19] and USA [20] also adopted this terminology, which was considered more general than the original “LIBS.”

The LIPS acronym is now deprecated, after the First International LIBS Conference (LIBS 2000), organized by the Pisa group in Tirrenia, Italy [21]. It is nevertheless curious that two of the major contributions to LIBS, which introduced two techniques still widely used nowadays, were in fact referring to LIPS as the name of the technique.

The first key paper was published in 1988 by the Sabsabi group in Canada and reported on an alternative experimental configuration in which the laser energy is delivered on the sample surface in two pulses, suitably delayed [22]. The authors reported a considerable intensity enhancement in the spectral signal which was substantially independent on the interpulse delay. The physical explanation of the enhancement in double-pulse configuration was fully explained only several years after by the Pallechi group in Pisa [23, 24], in terms of the reduced plasma shielding of the plasma produced by the second laser pulse, due to a reduction of the environmental gas density behind the shock wave produced by the first laser pulse. It is worth noting that the essential role of the first shock wave in double-pulse LIBS was firstly hypothesized by the Russian researcher Sergei Pershin; unfortunately, his research, published on a Russian journal [25], went generally unnoticed, and the author was credited for his original intuition only recently [26].

The second important paper was published by the Pisa group in 1999 and proposed a new procedure for standardless LIBS analysis called calibration-free LIPS (now known as CF-LIBS) [27]. Interestingly enough, an extended description of the method was published the same year on *Laser and Particle Beams* [28]. The Ciucci et al. paper is the most quoted research paper (thus excluding books and reviews) in the history of LIBS.

Many applications and improvements of the CF-LIBS method have been proposed since the original papers of 1999. One of the most useful procedures is the compensation




FIGURE 3: Schematic representation of the self-absorption effect, showing the balance between spontaneous emission and absorption (stimulated emission is normally negligible in LIBS plasmas).

of self-absorption effects in the LIBS plasma, which produce a non-linear dependence between the analyte concentration and the LIBS line intensity (see Figure 3).

The effect of self-absorption in laser-induced plasmas was studied in a key paper by the Winefordner group at the University of Florida in Gainesville, USA [29], but the implication of this research would not be transferred to CF-LIBS until the beginning of the XXI century.

#### 4. 2001–2021: XXI Century LIBS

The first proposal to use the curve-of-growth approach to compensate for self-absorption effects in LIBS plasmas was published by the Pisa group in 2002 [30]. A simple experimental method for evaluating the self-absorption effect and compensating it using a duplicating optical path mirror was proposed by the Gainesville group of Nicolò Omenetto in 2009 [31]. This method is conceptually very simple, although its realization is rather complex and limited to a laboratory setup. A more versatile method was proposed by the Paleschi group [32], as a generalization of the theoretical work of Amamou et al. [33]. According to this method, the degree of self-absorption of a given emission line can be calculated and, eventually, compensated, by measuring the intensity and full width at half maximum of the emission line and the measured plasma electron number density, once the Stark broadening coefficient of the line is known [34].

The method proposed by the Pisa group offers the possibility of measuring the plasma electron number density from the broadening of the Balmer alpha hydrogen line [35]. The problem was studied in 2013 by Pardini et al. [36].

Despite the fact that the Pisa method was initially developed for improving the predictions of the calibration-free LIBS technique, its applications have been extended to many situations in which the self-absorption effects are important (see [37] for a detailed discussion). Particularly important in this framework is the criticism to the “branching ratio” method for assessing the self-absorption effect of a spectral line published in 2021 by Urbina Medina et al. [38].

The ability of compensating for self-absorption effects opened new perspectives in the determination of spectroscopic fundamental parameters as transition probabilities [39, 40] and Stark broadening coefficients [41, 42].

One of the fundamental hypotheses for the application of the calibration-free LIBS is the fulfilment of the local thermal equilibrium (LTE) condition [43]. An important result, obtained at the end of the first decade of the century, was the extension of the McWhirter criterion for local thermal equilibrium to non-stationary and non-homogeneous LIBS plasmas [44]. The implication of that research confirmed the necessity, for the use of CF-LIBS, to use time-resolved detectors. A paper by Grifoni et al. in 2014 [45] provided a simple tool for extracting time-resolved information from time-integrated spectra, exploiting the differences between two or more spectra taken at different time delays.

The first decade of the century witnessed a great improvement in the performances of the LIBS technique, which accelerated its acceptance as a powerful analytical technique. In 2004, the Gainesville group published a paper, in which LIBS was defined as a possible future superstar among the atomic spectrometric techniques [46], and in 2010, David Hann and Nicolò Omenetto published an important review on Basic Diagnostics and Plasma-Particle Interactions [47] which at present time is the most quoted review paper in the LIBS history.

Many applications of LIBS in the analysis of biological materials [48–55], cultural heritage and archaeology [56–62], industry [63–65], and environment [66–69] have been reported.

The group of Javier Laserna at Malaga University, Spain, demonstrated the feasibility of performing stand-off LIBS analysis at long distances (>10 meters) using an open-path configuration [70, 71], thus extending dramatically the possible applications of LIBS for the analysis of industrial or environmental samples in hostile environment. Among the many exploitations of the LIBS technique, it is worth mentioning the results of a recent European project (LACOMORE—laser-based continuous monitoring and resolution of steel grades in sequence casting machines), aimed at the optimization of the continuous casting process of steel [72, 73]. The project involved the world’s two most active groups in LIBS development and research of Malaga and Pisa. In the framework of that project, an open-path double-pulse LIBS instrument was successfully used for remote analysis (about 6 meters) of steel up to 900 C temperature.

An important evolution of the LIBS technique, in the first decade of the century, was the introduction of ultra-short lasers sources for plasma generation [74–76]. Femtosecond laser pulses produce neater craters on the sample surface compared to nanosecond lasers, and the resulting spectra are characterized by a lower continuum emission because of the temporal separation between the ablation phenomenon and the creation of the plasma, which inhibits the laser-plasma interaction phenomena. The advantages of femtosecond LIBS were exploited for sub-micrometric in-depth measurements [77], analysis of biological tissues [78], and environmental applications [79]. Combined

nanosecond-femtosecond [80, 81] and femtosecond-femtosecond [82] dual-pulse analysis was also proposed. The possibility of maintaining the laser beam collimation at very long distances, due to the filamentation/self-focusing effects that characterize the propagation of fs-laser beams in atmosphere [79, 83–85], has triggered many innovative applications of stand-off LIBS, including the remote analysis of cultural heritage [86], biological materials [87], geological samples [88], and explosives [89, 90].

After the modest results of the proposals to use LIBS for Homeland Defense, in the years following the 9/11 tragic events [71, 91–94], LIBS regained some public consideration for its possible application in the field of space exploration [95]. Finally, on August 6, 2012, the NASA Curiosity rover landed on Mars, carrying a LIBS instrument [96] which is still operating after more than 10 years of activity, resulting in hundreds of thousands of spectra taken on Martian rocks. The data obtained by the LIBS instrument on Mars certainly helped in better understanding the Martian geology and contributed to the search of former life traces on the planet, opening the way to the use of LIBS in two other missions landed on Mars in 2021 (the NASA SuperCam instrument, mounted on the Perseverance rover [97], and the Chinese MarSCoDe mounted on the Zhurong Mars rover [98]). However, the success of LIBS on Mars benefitted more, if possible, the development of LIBS research on Earth. In fact, the first LIBS Mars mission spurred the development of compact hand-held LIBS spectrometers, which rapidly arrived on the market of scientific instrumentation for metal analysis [99, 100], nuclear industry [101], and environmental applications [102], to cite some of the most important applications.

The hand-held LIBS instruments represented an impressive advance with respect to the conventional laboratory or mobile LIBS instrumentation [103], but their compact size unavoidably imposes some compromise in the analytical performances of the instrumentation. The limitations of the hand-held LIBS hardware require the use of sophisticated chemometric techniques for extracting useful information from the spectra [99, 100, 104].

The use of advanced chemometric tools (artificial neural network, ANN) was firstly introduced in LIBS in 1998 [105]. However, these techniques became widely used only in the second decade of the century. The chemometric techniques can be used for simplification (for example, principal component analysis [106]), classification (self-organizing map (SOM) [107], support vector machine (SVM) [108], graph clustering (GC) [109], random forest (RF) [110], ANN [111], etc.), and quantification of the LIBS spectra (partial least squares (PLS) analysis [112], ANN [113], etc.). Numerous applications based on machine learning and chemometric analysis of LIBS spectra have been proposed in recent years. The most impressive are probably the applications to animal and human health (early detection of cancer, for example, [48]), but many other examples can be cited in cultural heritage [114], energy production [115], space exploration [116], and several other fields. LIBS elemental imaging [117, 118], which represents one of the most interesting developments of the technique, can produce

millions of spectra. The construction and interpretation of LIBS elemental maps also benefit the chemometric algorithms [119].

The growing complexity of the chemometric algorithms currently used in LIBS has stimulated some research aimed to obtain a better interpretability of the results obtained [120, 121]. An up-to-date description of the most advanced chemometric techniques can be found in [122].

Among the emerging new approaches to LIBS analysis, it is worth mentioning two interesting variations of LIBS which were proposed in the second decade of the century.

In 2011, Rick Russo proposed a technique for isotopic analysis by LIBS based on the detection of molecular emission in the spectrum. The isotopic shift of diatomic oxides or fluorides of the elements is typically larger than the ones of the corresponding atoms, thus allowing for an easier separation of the characteristic emission pattern of the different isotopes. The technique was called laser ablation molecular isotopic spectrometry (LAMIS) [123], and its effectiveness was demonstrated for the isotopic analysis of several interesting elements [124].

Another very interesting alternative approach to LIBS analysis was suggested by Alessandro De Giacomo and his group in Bari, Italy, in 2013. de Giacomo et al. reported on the use of nanoparticles to enhance the LIBS signal on metal targets [125]. The enhancement observed was comparable to the ones obtained in double-pulse LIBS, but obtained with a simpler, in principle, experimental apparatus. The mechanism of the enhancement in nanoparticle-enhanced LIBS (NELIBS) is very different from the one observed in DP-LIBS (total ablated mass, plasma electron density, and temperature are found very similar between NELIBS and conventional LIBS at the same energy). The mechanism hypothesized by the inventors of the technique involves a much larger atomization of the ablated mass in NELIBS, produced by the intense electric field which is created between one nanoparticle and the other. A model of the phenomenon has been recently published [126], which also explains the dependence of the enhancement on the distance between the nanoparticles (which in turn reflects in a strong dependence of the enhancement on the concentration of the deposited nanoparticle).

The full potential of NELIBS has probably not been yet explored, but this method can possibly change the conventional narrative of LIBS as a mediocre analytical technique. The applications of NELIBS should be mainly confined to the laboratory, though, because of the need of treating, although minimally, the sample under analysis. Proposals have been presented for putting together NELIBS and DP-LIBS, in a way to possibly combine the two enhancements [127]. Finally, in the analysis of insulators, NELIBS offers the advantage of enhancing the signal while reducing the surface damages, a feature that makes this approach particularly useful for the study of precious stones [128], for example.

The possibility of obtaining readable LIBS spectra from minimal quantity of ablated material was specifically studied by the Malaga group of Javier Laserna, which demonstrated the feasibility of LIBS for analysis of single nanoparticles

[129] and, in more recent papers [130–132], essentially debunked the standard narrative describing LIBS as a low sensitivity technique (in [130], a limit of detection of 60 attograms was demonstrated in the LIBS analysis of single copper nanoparticles).

## 5. Conclusions

The LIBS technique has made incredible progresses during its 40 years of existence. Certainly, the evolution of LIBS has been favoured by the technological evolution of the key instrumental parts (lasers, spectrometers, and detectors); however, the improved knowledge of the basic phenomena involved in the laser-sample, laser-plasma, and plasma-sample interactions has helped in better modelling the complex chemical and physical phenomena involved in the generation of the LIBS spectral signal. In this contribution, we have tried to retrace the history of LIBS to its fundamental papers, several of which were published in this journal. The list is far from being complete. A search on the Scopus® database with keywords (LIBS OR LIPS) reports, after removal of non-pertinent works, more than 12,000 papers published on the topic since the first two in 1981, with a growth rate in 2021 of more than 1,000 papers per year. Nevertheless, we hope to have given an idea of the distance that the technique has travelled from the first laboratory application to the present interplanetary missions.

There are still many steps to do and many obstacles to overcome before the LIBS technique could be considered to have growth to its full potential, but we have no doubts that some of these important steps will be presented and commented on this journal, as it was in the past with several key publications which still represent fundamental milestones in LIBS research.

## Data Availability

The relevant data are available from the author upon reasonable request.

## Conflicts of Interest

The author declares that there are no conflicts of interest regarding the publication of this paper.

## References

- [1] F. Brech and L. Cross, “Optical microemission stimulated by a ruby MASER,” *Applied Spectroscopy*, vol. 16, p. 59, 1962.
- [2] V. Palleschi, “Laser-induced breakdown spectroscopy: principles of the technique and future trends,” *ChemTexts*, vol. 6, no. 2, pp. 18–16, 2020.
- [3] E. Tognoni, V. Palleschi, M. Corsi, and G. Cristoforetti, “Quantitative micro-analysis by laser-induced breakdown spectroscopy: a review of the experimental approaches,” *Spectrochimica Acta, Part B: Atomic Spectroscopy*, vol. 57, no. 7, pp. 1115–1130, 2002.
- [4] A. Kramida, K. Olsen, and Y. Ralchemko, “NIST LIBS database,” 2022, <https://physics.nist.gov/PhysRefData/ASD/LIBS/lib-form.html>.
- [5] G. Amato, G. Cristoforetti, S. Legnaioli et al., “Progress towards an unassisted element identification from laser induced breakdown spectra with automatic ranking techniques inspired by text retrieval,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 65, no. 8, pp. 664–670, 2010.
- [6] L. J. Radziemski and T. R. Loree, “Laser-induced breakdown spectroscopy: time-resolved spectrochemical applications,” *Plasma Chemistry and Plasma Processing*, vol. 1, no. 3, pp. 281–293, 1981.
- [7] T. R. Loree and L. J. Radziemski, “Laser-induced breakdown spectroscopy: time-integrated applications,” *Plasma Chemistry and Plasma Processing*, vol. 1, no. 3, pp. 271–279, 1981.
- [8] L. J. Radziemski, T. R. Loree, D. A. Cremers, and N. M. Hoffman, “Time-resolved laser-induced breakdown Spectrometry of aerosols,” *Analytical Chemistry*, vol. 55, no. 8, pp. 1246–1252, 1983.
- [9] D. A. Cremers, L. J. Radziemski, and T. R. Loree, “Spectrochemical analysis of liquids using the laser spark,” *Applied Spectroscopy*, vol. 38, no. 5, pp. 721–729, 1984.
- [10] L. J. Radziemski, D. A. Cremers, and T. R. Loree, “Detection of beryllium by laser-induced-breakdown spectroscopy,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 38, no. 1-2, pp. 349–355, 1983.
- [11] J. A. Millard, R. H. Dalling, and L. J. Radziemski, “Time-resolved laser-induced breakdown Spectrometry for the rapid determination of beryllium in beryllium-copper alloys,” *Applied Spectroscopy*, vol. 40, no. 4, pp. 491–494, 2016.
- [12] J. R. Wachter and D. A. Cremers, “Determination of uranium in solution using laser-induced breakdown spectroscopy,” *Applied Spectroscopy*, vol. 41, no. 6, pp. 1042–1048, 2016.
- [13] M. Essien, L. J. Radziemski, and J. Sneddon, “Detection of cadmium, lead and zinc in aerosols by laser-induced breakdown Spectrometry,” *J Anal At Spectrom*, vol. 3, no. 7, pp. 985–988, 1988.
- [14] V. Palleschi, M. A. Harith, A. Salvetti, D. P. Singh, and M. Vaselli, “Time-resolved laser experiment for quantitative determination of pollutant concentrations in air,” *Laser and Particle Beams*, vol. 9, 1991.
- [15] C. Aragón, J. A. Aguilera, and J. Campos, “Determination of carbon content in molten steel using laser-induced breakdown spectroscopy,” *Applied Spectroscopy*, vol. 47, no. 5, pp. 606–608, 1993.
- [16] J. A. Aguilera, C. Aragón, and J. Campos, “Determination of carbon content in steel using laser-induced breakdown spectroscopy,” *Applied Spectroscopy*, vol. 46, no. 9, pp. 1382–1387, 1992.
- [17] M. Sabsabi and P. Cielo, “Quantitative analysis of aluminum alloys by laser-induced breakdown spectroscopy and plasma characterization,” *Applied Spectroscopy*, vol. 49, no. 4, pp. 499–507, 1995.
- [18] C. Haisch and U. Panne, “Laser-induced plasma spectroscopy (LIPS) in action,” *Spectroscopy Europe*, vol. 9, pp. 8–14, 1997.
- [19] L. St-Onge, M. Sabsabi, and P. Cielo, “Quantitative analysis of additives in solid zinc alloys by laser-induced plasma Spectrometry,” *J Anal At Spectrom*, vol. 12, no. 9, pp. 997–1004, 1997.
- [20] M. Martin and M.-D. Cheng, “Detection of chromium aerosol using time-resolved laser-induced plasma spectroscopy,” *Applied Spectroscopy*, vol. 54, no. 9, pp. 1279–1285, 2000.
- [21] V. Palleschi, “The first international conference on laser-induced plasma spectroscopy and applications (LIBS 2000),”

- Spectrochimica Acta Part B Atomic Spectroscopy*, vol. 56, 2001.
- [22] L. St-Onge, M. Sabsabi, and P. Cielo, "Analysis of solids using laser-induced plasma spectroscopy in double-pulse mode," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 53, no. 3, pp. 407–415, 1998.
- [23] M. Corsi, G. Cristoforetti, M. Giuffrida et al., "Three-dimensional analysis of laser induced plasmas in single and double pulse configuration," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 59, no. 5, pp. 723–735, 2004.
- [24] S.-B. Wen, X. Mao, R. E. Russo et al. "Comment on "three-dimensional analysis of laser induced plasmas in single and double pulse configuration"," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 60, no. 6, pp. 870–872, 2005.
- [25] S. Pershin, "Transformation of the optical spectrum of a laser plasma when a surface is irradiated with a double pulse," *Soviet Journal of Quantum Electronics*, vol. 19, no. 2, pp. 215–218, 1989.
- [26] G. Cristoforetti and V. Palleschi, *Double-Pulse Laser Ablation of Solid Targets in Ambient Gas: Mechanisms and Effects*, Nova Science Publishers, Inc, Hauppauge, NY, USA, 2011.
- [27] A. Ciucci, M. Corsi, V. Palleschi, S. Rastelli, A. Salvetti, and E. Tognoni, "New procedure for quantitative elemental analysis by laser-induced plasma spectroscopy," *Applied Spectroscopy*, vol. 53, no. 8, pp. 960–964, 1999.
- [28] A. Ciucci, V. Palleschi, S. Rastelli, A. Salvetti, D. P. Singh, and E. C. F.-L. I. P. S. Tognoni, "CF-LIPS: a new approach to LIPS spectra analysis," *Laser and Particle Beams*, vol. 17, no. 4, pp. 793–797, 1999.
- [29] I. B. B. Gornushkin, J. M. M. Anzano, L. A. A. King, B. W. W. Smith, N. Omenetto, and J. D. D. Winefordner, "Curve of growth methodology applied to laser-induced plasma emission spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 54, no. 3-4, pp. 491–503, 1999.
- [30] D. Bulajic, M. Corsi, G. Cristoforetti et al., "A procedure for correcting self-absorption in calibration free-laser induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 57, no. 2, pp. 339–353, 2002.
- [31] H.-Y. Moon, K. K. Herrera, N. Omenetto, B. W. Smith, and J. D. Winefordner, "On the usefulness of a duplicating mirror to evaluate self-absorption effects in laser induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 64, no. 7, pp. 702–713, 2009.
- [32] A. M. El Sherbini, T. M. El Sherbini, H. Hegazy et al., "Evaluation of self-absorption coefficients of aluminum emission lines in laser-induced breakdown spectroscopy measurements," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 60, no. 12, pp. 1573–1579, 2005.
- [33] H. Amamou, A. Bois, B. Ferhat, R. Redon, B. Rossetto, and P. Matheron, "Correction of self-absorption spectral line and ratios of transition probabilities for homogeneous and LTE plasma," *Journal of Quantitative Spectroscopy and Radiative Transfer*, vol. 75, no. 6, pp. 747–763, 2002.
- [34] H. R. Griem, "Stark broadening," *Advances in Atomic and Molecular Physics*, vol. 11, pp. 331–359, 1976.
- [35] M. A. Gigosos, M. Á. González, and V. Cardenoso, "Computer simulated balmer-alpha, beta and gamma Stark line profiles for non-equilibrium plasmas Diagnostics," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 58, no. 8, pp. 1489–1504, 2003.
- [36] L. Pardini, S. Legnaioli, G. Lorenzetti et al., "On the determination of plasma electron number density from Stark broadened hydrogen balmer series lines in laser-induced breakdown spectroscopy experiments," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 88, pp. 98–103, 2013.
- [37] F. Rezaei, G. Cristoforetti, E. Tognoni, S. Legnaioli, V. Palleschi, and A. Safi, "A review of the current analytical approaches for evaluating, compensating and exploiting self-absorption in laser induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 169, Article ID 105878, 2020.
- [38] I. A. Urbina Medina, D. D. Carneiro, S. Rocha, E. E. Farias, F. O. Bredice, and V. Palleschi, "Branching ratio method for assessing optically thin conditions in laser-induced plasmas," *Applied Spectroscopy*, vol. 75, pp. 000370282110067–000370282110780, 2021.
- [39] I. Urbina, D. Carneiro, S. Rocha, E. Farias, F. Bredice, and V. Palleschi, "Measurement of atomic transition probabilities with laser-induced breakdown spectroscopy using the 3D Boltzmann plot method," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 154, pp. 91–96, 2019.
- [40] J. A. Aguilera, C. Aragón, and J. Manrique, "Method for measurement of transition probabilities by laser-induced breakdown spectroscopy based on CSigma graphs—application to Ca II spectral lines," *Journal of Quantitative Spectroscopy and Radiative Transfer*, vol. 160, pp. 10–18, 2015.
- [41] F. Poggialini, B. Campanella, R. Jafer et al., "Determination of the Stark broadening coefficients of tantalum emission lines by time-independent extended C-sigma method," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 167, Article ID 105829, 2020.
- [42] F. Bredice, F. O. O. Borges, H. Sobral et al., "Measurement of Stark broadening of MnI and MnII spectral lines in plasmas used for laser-induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 62, no. 11, pp. 1237–1245, 2007.
- [43] H. R. Griem, "Validity of local thermal Equilibrium in plasma spectroscopy," *Physical Review A*, vol. 131, no. 3, pp. 1170–1176, 1963.
- [44] G. Cristoforetti, A. de Giacomo, M. Dell'Aglio et al., "Local thermodynamic Equilibrium in laser-induced breakdown spectroscopy: beyond the McWhirter criterion," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 65, no. 1, pp. 86–95, 2010.
- [45] E. Grifoni, S. Legnaioli, M. Lezzerini, G. Lorenzetti, S. Pagnotta, and V. Palleschi, "Extracting time-resolved information from time-integrated laser-induced breakdown spectra," *Journal of Spectroscopy*, vol. 2014, Article ID 849310, 5 pages, 2014.
- [46] J. D. Winefordner, I. B. Gornushkin, T. Correll, E. Gibb, B. W. Smith, and N. Omenetto, "Comparing several atomic spectrometric methods to the super stars: special emphasis on laser induced breakdown Spectrometry, LIBS, a future super star," *J Anal At Spectrom*, vol. 19, no. 9, pp. 1061–1083, 2004.
- [47] D. W. Hahn and N. Omenetto, "Laser-induced breakdown spectroscopy (LIBS), Part I: review of basic Diagnostics and plasma—particle interactions: still-challenging issues within the analytical plasma community," *Applied Spectroscopy*, vol. 64, no. 12, pp. 335A–336A, 2010.
- [48] R. Gaudiuso, N. Melikechi, Z. A. Abdel-Salam et al., "Laser-induced breakdown spectroscopy for human and animal health: a review," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 152, pp. 123–148, 2019.
- [49] M. Corsi, G. Cristoforetti, M. Hidalgo et al., "Application of laser-induced breakdown spectroscopy technique to hair

- tissue mineral analysis," *Applied Optics*, vol. 42, no. 30, p. 6133, 2003.
- [50] R. A. Multari, D. A. Cremers, and M. L. Bostian, "Use of laser-induced breakdown spectroscopy for the differentiation of pathogens and viruses on substrates," *Applied Optics*, vol. 51, no. 7, pp. B57–B64, 2012.
- [51] R. A. Multari, D. A. Cremers, T. Scott, and P. Kendrick, "Detection of pesticides and dioxins in tissue fats and rendering oils using laser-induced breakdown spectroscopy (LIBS)," *Journal of Agricultural and Food Chemistry*, vol. 61, no. 10, pp. 2348–2357, 2013.
- [52] R. A. Multari, D. A. Cremers, J. M. Dupre, and J. E. Gustafson, "The use of laser-induced breakdown spectroscopy for distinguishing between bacterial pathogen species and strains," *Applied Spectroscopy*, vol. 64, no. 7, pp. 750–759, 2010.
- [53] R. A. Multari, D. A. Cremers, J. A. M. Dupre, and J. E. Gustafson, "Detection of biological contaminants on foods and food surfaces using laser-induced breakdown spectroscopy (LIBS)," *Journal of Agricultural and Food Chemistry*, vol. 61, no. 36, pp. 8687–8694, 2013.
- [54] S. J. Rehse, "Biomedical applications of LIBS," *Springer Series in Optical Sciences*, vol. 182, pp. 457–488, 2014.
- [55] S. J. Rehse, H. Salimnia, and A. W. Miziolek, "Laser-induced breakdown spectroscopy (LIBS): an overview of recent progress and future potential for biomedical applications," *Journal of Medical Engineering and Technology*, vol. 36, no. 2, pp. 77–89, 2012.
- [56] J. H. Scholten, J. M. Teule, V. Zafirooulos, and R. M. A. Heeren, "Controlled laser cleaning of painted artworks using accurate beam manipulation and on-line LIBS-detection," *Journal of Cultural Heritage*, vol. 1, pp. S215–S220, 2000.
- [57] S. Klein, J. Hildenhausen, K. Dickmann, T. Stratoudaki, and V. Zafirooulos, "LIBS-spectroscopy for monitoring and control of the laser cleaning process of stone and medieval glass," *Journal of Cultural Heritage*, vol. 1, pp. S287–S292, 2000.
- [58] D. Anglos and V. Detalle, "Cultural heritage applications of LIBS," in *Laser-Induced Breakdown Spectroscopy*, pp. 531–554, Springer, Berlin, Germany, 2014.
- [59] M. Simileanu, "Libs quantitative analyses of bronze objects for cultural heritage applications," *Romanian Reports in Physics*, vol. 68, pp. 203–209, 2016.
- [60] L. M. Osorio, L. V. P. Cabrera, M. A. A. Garcia, T. F. Reyes, and I. Ravelo, "Portable LIBS system for determining the composition of multilayer structures on objects of cultural value," *J Phys Conf Ser*, vol. 274, Article ID 012093, 2011.
- [61] G. Cristoforetti, S. Legnaioli, V. Palleschi, L. Pardini, A. Salvetti, and E. M. Tognoni, "A new mobile instrument for in situ standard-less LIBS analysis of cultural heritage," *Proceedings of the Proceedings of SPIE The International Society for Optical Engineering*, vol. 5857, 2005.
- [62] A. Arafat, M. Na'as, V. Kantarelou et al., "Combined in situ micro-XRF, LIBS and SEM-EDS analysis of base metal and corrosion products for islamic copper alloyed artefacts from umm qais museum, Jordan," *Journal of Cultural Heritage*, vol. 14, no. 3, pp. 261–269, 2013.
- [63] D. Bulajic, G. Cristoforetti, M. Corsi et al., "Diagnostics of high-temperature steel pipes in industrial environment by laser-induced breakdown spectroscopy technique: the LIBSGRAIN project," *Spectrochimica Acta, Part B: Atomic Spectroscopy*, vol. 57, no. 7, pp. 1181–1192, 2002.
- [64] R. Noll, V. Sturm, M. Stepputat, A. Whitehouse, J. Young, and P. Evans, "Industrial applications of LIBS," in *Laser-Induced Breakdown Spectroscopy (LIBS)*, A. W. Miziolek, V. Palleschi, and I. Schechter, Eds., pp. 400–439, Cambridge University Press, Cambridge, UK, 2006.
- [65] C. Li, C. L. Feng, H. Y. Oderji, G. N. Luo, and H. Ding, "Review of LIBS application in nuclear fusion technology," *Frontiers in Physiology*, vol. 11, no. 6, Article ID 114214, 2016.
- [66] T. Karpate, R. Nayak, and S. Libs, *A Potential Tool for Industrial/Agricultural Waste Water Analysis*, Harvard University, Cambridge, MA, USA, 2016.
- [67] A. M. Popov, F. Colao, and R. Fantoni, "Spatial confinement of laser-induced plasma to enhance LIBS sensitivity for trace elements determination in soils," *J Anal At Spectrom*, vol. 25, no. 6, p. 837, 2010.
- [68] G. S. Senesi, M. Dell'Aglio, R. Gaudiuso et al., "Heavy metal concentrations in soils as determined by laser-induced breakdown spectroscopy (LIBS), with special emphasis on chromium," *Environmental Research*, vol. 109, no. 4, pp. 413–420, 2009.
- [69] L. C. Nunes, G. G. A. De Carvalho, D. Santos, and F. J. Krug, "Determination of Cd, Cr and Pb in phosphate fertilizers by laser-induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 97, pp. 42–48, 2014.
- [70] S. Palanco and J. Laserna, "Remote sensing instrument for solid samples based on open-path atomic emission Spectrometry," *Review of Scientific Instruments*, vol. 75, no. 6, pp. 2068–2074, 2004.
- [71] F. J. Fortes and J. J. Laserna, "The development of fieldable laser-induced breakdown spectrometer: No limits on the horizon," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 65, no. 12, pp. 975–990, 2010.
- [72] G. Lorenzetti, S. Legnaioli, E. Grifoni, S. Pagnotta, and V. Palleschi, "Laser-based continuous monitoring and resolution of steel grades in sequence casting machines," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 112, pp. 1–5, 2015.
- [73] L. M. Cabalín, T. Delgado, J. Ruiz, D. Mier, and J. J. Laserna, "Stand-off laser-induced breakdown spectroscopy for steel-grade intermix detection in sequence casting operations. At-line monitoring of temporal evolution versus predicted mathematical model," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 146, pp. 93–100, 2018.
- [74] B. Le Drogoff, J. Margot, M. Chaker et al., "Temporal characterization of femtosecond laser pulses induced plasma for spectrochemical analysis of aluminum alloys," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 56, no. 6, pp. 987–1002, 2001.
- [75] K. L. Eland, D. N. Stratis, D. M. Gold, S. R. Goode, and S. M. Angel, "Energy dependence of emission intensity and temperature in a LIBS plasma using femtosecond excitation," *Applied Spectroscopy*, vol. 55, no. 3, pp. 286–291, 2001.
- [76] V. Margetic, K. Niemax, and R. Hergenröder, "A study of non-linear calibration graphs for brass with femtosecond laser-induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 56, no. 6, pp. 1003–1010, 2001.
- [77] V. Margetic, M. Bolshov, A. Stockhaus, K. Niemax, and R. Hergenröder, "Depth profiling of multi-layer samples using femtosecond laser ablation," *J Anal At Spectrom*, vol. 16, no. 6, pp. 616–621, 2001.
- [78] Y. Moon, J. H. Han, J. Choi, S. Shin, Y.-C. Kim, and S. Jeong, "Mapping of cutaneous melanoma by femtosecond laser-

- induced breakdown spectroscopy,” *Journal of Biomedical Optics*, vol. 24, no. 03, p. 1, 2018.
- [79] M. Hu, S. Shi, M. Yan, E. Wu, and H. Zeng, “Femtosecond laser-induced breakdown spectroscopy by multidimensional plasma grating,” *J Anal At Spectrom*, vol. 37, no. 4, pp. 841–848, 2022.
- [80] J. Scaffidi, J. Pender, W. Pearman et al., “Dual-pulse laser-induced breakdown spectroscopy with combinations of femtosecond and nanosecond laser pulses,” *Applied Optics*, vol. 42, no. 30, p. 6099, 2003.
- [81] V. I. Babushok, F. C. DeLucia, J. L. Gottfried, C. A. Munson, and A. W. Miziolek, “Double pulse laser ablation and plasma: laser induced breakdown spectroscopy signal enhancement,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 61, no. 9, pp. 999–1014, 2006.
- [82] V. Piñon, C. Fotakis, G. Nicolas, and D. Anglos, “Double pulse laser-induced breakdown spectroscopy with femtosecond laser pulses,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 63, no. 10, pp. 1006–1010, 2008.
- [83] K. Stelmazczyk, P. Rohwetter, G. Méjean et al., “Long-distance remote laser-induced breakdown spectroscopy using filamentation in air,” *Applied Physics Letters*, vol. 85, no. 18, pp. 3977–3979, 2004.
- [84] P. Polynkin, M. Burger, I. Jovanovic, I. Jovanovic, and I. Jovanovic, “Filament-induced breakdown spectroscopy with structured beams,” *Optics Express*, vol. 28, no. 24, pp. 36812–36821, 2020.
- [85] M. Hu, S. Xu, S. Yuan, and H. Zeng, “Breakdown spectroscopy induced by nonlinear interactions of femtosecond laser filaments and multidimensional plasma gratings,” *Ultrafast Science*, vol. 3, 2023.
- [86] S. Tzortzakis, D. Anglos, and D. Gray, “Ultraviolet laser filaments for remote laser-induced breakdown spectroscopy (LIBS) analysis: applications in cultural heritage monitoring,” *Optics Letters*, vol. 31, no. 8, pp. 1139–1141, 2006.
- [87] H. L. Xu, G. Méjean, W. Liu et al., “Remote detection of similar biological materials using femtosecond filament-induced breakdown spectroscopy,” *Applied Physics B: Photophysics and Laser Chemistry*, vol. 87, no. 1, pp. 151–156, 2007.
- [88] S. Abdul Kalam, S. V. Balaji Manasa Rao, M. Jayananda, and S. Venugopal Rao, “Standoff femtosecond filament-induced breakdown spectroscopy for classification of geological materials,” *J Anal At Spectrom*, vol. 35, no. 12, pp. 3007–3020, 2020.
- [89] A. K. Shaik, N. R. Epuru, H. Syed et al. “Femtosecond laser induced breakdown spectroscopy based standoff detection of explosives and discrimination using principal component analysis,” *Optics Express*, vol. 26, no. 7, pp. 8069–8083, 2018.
- [90] A. K. Shaik and V. R. Soma, “Standoff discrimination and trace detection of explosive molecules using femtosecond filament induced breakdown spectroscopy combined with silver nanoparticles,” *OSA Continuum*, vol. 2, no. 3, pp. 554–562, 2019.
- [91] R. González, P. Lucena, L. M. Tobaría, and J. J. Laserna, “Standoff LIBS detection of explosive residues behind a barrier,” *J Anal At Spectrom*, vol. 24, no. 8, p. 1123, 2009.
- [92] P. Lucena, A. Doña, L. M. Tobaría, and J. J. Laserna, “New challenges and insights in the detection and spectral identification of organic explosives by laser induced breakdown spectroscopy,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 66, no. 1, pp. 12–20, 2011.
- [93] F. C. DeLucia, A. C. Samuels, R. S. Harmon et al., “Laser-induced breakdown spectroscopy (LIBS): a promising versatile chemical sensor technology for hazardous material detection,” *IEEE Sensors Journal*, vol. 5, no. 4, pp. 681–689, 2005.
- [94] J. L. Gottfried, F. C. De Lucia, C. A. Munson, and A. W. Miziolek, “Standoff detection of chemical and biological threats using laser-induced breakdown spectroscopy,” *Applied Spectroscopy*, vol. 62, no. 4, pp. 353–363, 2008.
- [95] A. K. Knight, N. L. Scherbarth, D. A. Cremers, and M. J. Ferris, “Characterization of laser-induced breakdown spectroscopy (LIBS) for application to space exploration,” *Applied Spectroscopy*, vol. 54, no. 3, pp. 331–340, 2000.
- [96] R. C. Wiens, S. Maurice, B. Barraclough et al., “The ChemCam instrument suite on the Mars science laboratory (MSL) rover: body unit and combined system tests,” *Space Science Reviews*, vol. 170, no. 1-4, pp. 167–227, 2012.
- [97] R. C. Wiens, S. Maurice, S. H. Robinson et al., “The SuperCam instrument suite on the NASA Mars 2020 rover: body unit and combined system tests,” *Space Science Reviews*, vol. 217, pp. 1–87, 2021.
- [98] W. Xu, X. Liu, Z. Yan et al., “The MarSCoDe instrument suite on the Mars rover of China’s tianwen-1 mission,” *Space Science Reviews*, vol. 217, pp. 1–58, 2021.
- [99] M. S. Afgan, Z. Hou, and Z. Wang, “Quantitative analysis of common elements in steel using a handheld  $\mu$ -LIBS instrument,” *J Anal At Spectrom*, vol. 32, no. 10, pp. 1905–1915, 2017.
- [100] F. Poggialini, B. Campanella, S. Legnaioli, S. Pagnotta, S. Raneri, and V. Palleschi, “Improvement of the performances of a commercial hand-held laser-induced breakdown spectroscopy instrument for steel analysis using multiple artificial neural Networks,” *Review of Scientific Instruments*, vol. 91, no. 7, Article ID 073111, 2020.
- [101] E. Garlea, B. N. Bennett, M. Z. Martin, R. L. Bridges, G. L. Powell, and J. H. Leckey, “Novel use of a hand-held laser induced breakdown spectroscopy instrument to monitor hydride corrosion in uranium,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 159, Article ID 105651, 2019.
- [102] H. VerMeulen, J. Clausen, A. Mosell, M. Morgan, K. Messan, and S. Beal, “Application of laser induced breakdown (LIBS) for environmental, chemical, and biological sensing,” *Advanced Environmental, Chemical, and Biological Sensing Technologies XV*, pp. 68–90, Article ID 11007, 2019.
- [103] G. S. Senesi, R. S. Harmon, and R. R. Hark, “Field-portable and handheld laser-induced breakdown spectroscopy: historical review, current status and future prospects,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 175, Article ID 106013, 2021.
- [104] L. Zhou, Y. Luo, Y. Q. Zheng, and C. G. Huang, “Field determination of chromium, nickel, manganese, copper and silicon in steel product by hand-held laser induced breakdown spectrometer,” *Yejin Fenxi/Metallurgical Analysis*, vol. 39, pp. 8–12, 2019.
- [105] R. Sattmann, I. Mönch, H. Krause et al., “Laser-induced breakdown spectroscopy for polymer identification,” *Applied Spectroscopy*, vol. 52, no. 3, pp. 456–461, 1998.
- [106] S. Moncayo, L. Duponchel, N. Mousavipak et al., “Exploration of megapixel hyperspectral LIBS images using principal component analysis,” *J Anal At Spectrom*, vol. 33, no. 2, pp. 210–220, 2018.
- [107] S. Pagnotta, E. Grifoni, S. Legnaioli, M. Lezznerini, G. Lorenzetti, and V. Palleschi, “Comparison of brass alloys composition by laser-induced breakdown spectroscopy and self-organizing maps,” *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 103-104, pp. 70–75, 2015.



- [108] N. C. Dingari, I. Barman, A. K. Myakalwar, S. P. Tewari, and M. Kumar Gundawar, "Incorporation of Support vector machines in the LIBS toolbox for sensitive and robust classification amidst unexpected sample and system variability," *Analytical Chemistry*, vol. 84, no. 6, pp. 2686–2694, 2012.
- [109] E. Grifoni, S. Legnaioli, G. Lorenzetti, S. Pagnotta, and V. Palleschi, "Application of Graph theory to unsupervised classification of materials by laser-induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 118, pp. 40–44, 2016.
- [110] L. Sheng, T. Zhang, G. Niu et al., "Classification of iron ores by laser-induced breakdown spectroscopy (LIBS) combined with random forest (RF)," *J Anal At Spectrom*, vol. 30, no. 2, pp. 453–458, 2015.
- [111] D. Pokrajac, T. Vance, A. Lazarević et al., "Performance of multilayer perceptrons for classification of LIBS protein spectra," in *Proceedings of the 10th Symposium on Neural Network Applications in Electrical Engineering*, pp. 171–174, Belgrade, Serbia, September 2010.
- [112] P. Yaroshchuk, D. L. Death, and S. J. Spencer, "Comparison of principal components regression, partial Least Squares regression, multi-block partial Least Squares regression, and serial partial Least Squares regression algorithms for the analysis of Fe in iron ore using LIBS," *J. Anal. At. Spectrom*, vol. 27, no. 1, pp. 92–98, 2012.
- [113] J. B. Sirven, B. Bousquet, L. Canioni et al., "Qualitative and quantitative investigation of chromium-polluted soils by laser-induced breakdown spectroscopy combined with neural Networks analysis," *Analytical and Bioanalytical Chemistry*, vol. 385, no. 2, pp. 256–262, 2006.
- [114] S. Duchene, V. Detalle, R. Bruder, and J. Sirven, "Chemometrics and laser induced breakdown spectroscopy (LIBS) analyses for identification of wall paintings pigments," *Current Analytical Chemistry*, vol. 6, no. 1, pp. 60–65, 2010.
- [115] S. Legnaioli, B. Campanella, S. Pagnotta, F. Poggialini, and V. Palleschi, "Determination of ash content of coal by laser-induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 155, pp. 123–126, 2019.
- [116] J. B. Sirven, B. Sallé, P. Mauchien, J. L. Lacour, S. Maurice, and G. Manhès, "Feasibility study of rock identification at the surface of Mars by remote laser-induced breakdown spectroscopy and three chemometric methods," *J Anal At Spectrom*, vol. 22, no. 12, pp. 1471–1480, 2007.
- [117] L. Sancey, V. Motto-Ros, B. Busser et al., "Laser Spectrometry for multi-elemental imaging of biological tissues," *Scientific Reports*, vol. 4, no. 1, p. 6065, 2014.
- [118] B. Busser, S. Moncayo, J. L. Coll, L. Sancey, and V. Motto-Ros, "Elemental imaging using laser-induced breakdown spectroscopy: a new and promising approach for biological and medical applications," *Coordination Chemistry Reviews*, vol. 358, pp. 70–79, 2018.
- [119] L. Jolivet, M. Leprince, S. Moncayo, L. Sorbier, C.-P. Lienemann, and V. Motto-Ros, "Review of the recent advances and applications of LIBS-based imaging," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 151, pp. 41–53, 2019.
- [120] A. Safi, B. Campanella, E. Grifoni et al., "Multivariate calibration in laser-induced breakdown spectroscopy quantitative analysis: the dangers of a 'black box' approach and how to avoid them," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 144, pp. 46–54, 2018.
- [121] E. Képeš, J. Vrábek, O. Adamovsky et al., "Interpreting Support vector machines applied in laser-induced breakdown spectroscopy," *Analytica Chimica Acta*, vol. 1192, Article ID 339352, 2022.
- [122] V. Palleschi, *Chemometrics and Numerical Methods in LIBS*, John Wiley, Hoboken, NY, USA, 2022.
- [123] R. E. Russo, A. A. Bol'shakov, X. Mao, C. P. McKay, D. L. Perry, and O. Sorkhabi, "Laser ablation molecular isotopic Spectrometry," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 66, no. 2, pp. 99–104, 2011.
- [124] A. A. Bol'shakov, X. Mao, J. J. González, and R. E. Russo, "Laser ablation molecular isotopic Spectrometry (LAMIS): current state of the art," *J Anal At Spectrom*, vol. 31, no. 1, pp. 119–134, 2016.
- [125] A. de Giacomo, R. Gaudiuso, C. Koral, M. Dell'Aglio, and O. de Pascale, "Nanoparticle-enhanced laser-induced breakdown spectroscopy of metallic samples," *Analytical Chemistry*, vol. 85, no. 21, pp. 10180–10187, 2013.
- [126] A. de Giacomo, Z. Salajkova, and M. Dell'aglio, "A quantum chemistry approach based on the analogy with  $\pi$ -system in polymers for a rapid estimation of the resonance wavelength of nanoparticle systems," *Nanomaterials*, vol. 9, no. 7, pp. 929–1016, 2019.
- [127] F. Poggialini, B. Campanella, S. Legnaioli, S. Pagnotta, and V. Palleschi, "Investigating double pulse nanoparticle-enhanced laser-induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 167, Article ID 105845, 2020.
- [128] C. Koral, M. Dell'Aglio, R. Gaudiuso, R. Alrifai, M. Torelli, and A. de Giacomo, "Nanoparticle-enhanced laser induced breakdown spectroscopy for the noninvasive analysis of transparent samples and gemstones," *Talanta*, vol. 182, pp. 253–258, 2018.
- [129] F. J. Fortes, A. Fernández-Bravo, and J. Javier Laserna, "Chemical characterization of single micro- and nanoparticles by optical catapulting-optical trapping-laser-induced breakdown spectroscopy," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 100, pp. 78–85, 2014.
- [130] P. Purohit, F. J. Fortes, and J. J. Laserna, "Spectral identification in the attogram regime through laser-induced emission of single optically trapped nanoparticles in air," *Angewandte Chemie International Edition*, vol. 56, no. 45, pp. 14178–14182, 2017.
- [131] P. Purohit, F. J. Fortes, and J. J. Laserna, "Subfemtogram simultaneous elemental detection in multicomponent nanomatrices using laser-induced plasma emission spectroscopy within atmospheric pressure optical traps," *Analytical Chemistry*, vol. 91, no. 11, pp. 7444–7449, 2019.
- [132] P. Purohit, F. J. Fortes, and J. J. Laserna, "Atomization efficiency and photon yield in laser-induced breakdown spectroscopy analysis of single nanoparticles in an optical trap," *Spectrochimica Acta Part B: Atomic Spectroscopy*, vol. 130, pp. 75–81, 2017.